

HSCC/PG-III/SIC/CSSD/2010

Dated: 25.3.2010

To,

Sub : Tender for Out sourcing of CSSD Services for Sports Injury Centre, at Safdarjung Hospital New Delhi.- Clarifications of Queries of Pre Bid Meeting dated 11.3.2010 regarding

Bid Reference:: IFB. No. SIC-SJH/PG-III/01/2010/CSSD Date 2.3. 2010

Dear Sir,

With reference to the subject mentioned above, kindly find enclosed herewith the clarifications of the queries raised during the pre bid meeting held on 11.3.2010.

This document shall be treated as the part of tender documents. You are requested to submit all the amendments along with clarifications issued duly signed along with the tender documents.

Thanking You,

Yours Faithfully

(Anil Kumar)
General Manager(PG-III)

Dated: 25.3.2010

IFB. No. SIC-SJH/PG-III/01/2010/CSSD Date 2.3. 2010

**Sub : Tender for Out sourcing of CSSD Services for Sports Injury Centre,
Safdarjung Hospital New Delhi.**

Clarifications of Queries of Pre Bid Meeting dated 11.3.2010

	Bidder No.1	
Sl no	QUERY FROM THE BIDDER	REPLY FROM THE HOSPITAL
1	Dressing pads supply shall be whose responsibility and what types of containers shall be used .	This shall be responsibility of bidder
2	Indicators and integrations shall be provided by supplier and spours shall be provided by the hospital laboratory and tested there only.	All required consumables including biological indicators and equipments etc shall be the responsibility of bidder.
3	Packaging shall be double medical packaging in trays / baskets.	Packaging used shall be of medical graded paper with double layered for all trays and transportation means shall be baskets
4	As per requirement the space is inefficient, as separate R-O plant has not been incorporated in the drawing .Hence separate locations is to be specified from where the pipe line is to be connected . Please specify the space .	Hospital shall provide for soft water at one point with in the CSSD space indicated in the bid document. Bidder are required to satisfy with the quality of water available and specification for proposed RO plant in the sports Injury Centre before submission of bid. Any requirement beyond this provision shall be the responsibility of the bidder. The space indicated in the bid document shall be the final space to be provided by the Hospital.
5	Clarify the total CSSD area in sq.ft	The area of CSSD is approx 1280 sq feet.
6	The load for sterilization shall be only for sports injury hospital not for any attached unit.Cap. 50 bedded	The load under this tender is for Sports Injury Centre only.
7	E.O gas sterilization/ Plazma shall be beyond CSSD scope.	This is in the scope of hospital.
8	No of power points , fans , lights shall be fixed or left to suppliers discretions.	All is in the scope of bidder
9	Drain sump /outlet is to be provided in the CSSD where are internal drains shall be connected .	The one drain outlet from CSSD shall be provided by the hospital. Internal drainage shall be the responsibility of Bidder
10	No of Drums and Pas box shall be as per the CSSD drawing .	Yes drums and pass box shall be provided by bidder, however the actual requirement shall be as proposed by bidder to meet the total requirement

11	Stand by arrangements for sterilizer if required shall be of an Indian make as discussed & accepted in Pre bid meeting,	This shall be as clarified in Minutes of pre bid Meeting already uploaded to web site.
12	Linen to hospital account.	This will to Hospital account
13	Post sterilization record for 1 year only.	Bidders are required to maintain the records for minimum of two years.
14	Gloves, shoe covers and caps mask etc shall be provided by the hospital.	This shall be provided by hospital.
15	As sealing of medical Packaging is too much therefore Auto seaers should be atleast 3 in numbers in CSSD instead of one.	This shall be minimum of three auto sealers.
16	Closed transport Trolley should be 4-6 .	The total closed trolley shall be five in number.
17	Internal two shelf Trolley for transporting goods shall be- 2 to 3 .	Internal two shelf Trolley for transporting goods shall be- 3 (three)
18	Day to day housekeeping to hospital account.	Housekeeping within the CSSD area shall be bidders account .The disposal of refuse beyond CSSD area and from the hospital area shall be hospital responsibility.
	Bidder No. 2	
1	Clause No.10: Bid Prices/Sub clause 10.4 (Bidder should quote Man and material) : We seek clarification :what authorities want to convey by stating that charge as one time investment and how and when this will be disbursed to the successful bidder.	The said clause no 10.4 page 10 of bid document stands deleted. The bids are invited as per the provisions of the bid documents and relevant item of work under BOQ.
2	P.No.20 Terms and conditions of Outsourcing of CSSD services Clause 1.1 General Condition Sub clause 1.1.6 (The license agreement All concerned / and page no.30 contract termination clause sub clause 1.12.1 (Sport Injury Safdarjang ... New Delhi) : We seek clarification in case of termination of the contract by either parties, what would be the exit process and financial implications.	The bidders are required to submit a scheme of exit from the contract in case of termination. The bidder are required to quote the exit cost bidder would like to charge from the hospital as on whereas basis in the event of termination for every sixth month for the complete 60 months. The exit cost in between the periods shall on prorated basis.
3	Refer to the core objective of the tender. It has been observed that certified international standard to operate check and maintain record of each and every cycle of sterilisation process and which is reproducible as and when required : tendering authorities seems to be silent on this subject. The tender is more concerned about the equipment to be placed by the successful bidder for use. Why the tendering authority is asking for the imported equipment only when majority (more than 99%) of the equipment used in CSSD are indigenous both in Govt./private sector	The specification and origin of equipment shall be as per bid document supplied and additional information provided through web site. The record of CSSD services have to be maintained for a minimum period of two years. The level of quality services shall be as per 1.1.11 page 21 and other provisions of bid document.
4	Circulation of minutes of pre-bid meeting.	Up loaded on to the web site