

**PRE BID MEETING QUERY FROM BIDDERS AND HOSPITAL  
REPLY**

( SIGNED COPY OF THIS REPLY MUST BE INCLUDED ALONG WITH THE OFFER BY THE BIDDER )

**Clarification-2**

| <b>BIDDER -1</b> |   |   |
|------------------|---|---|
| <b>S.No.</b>     | <b>QUERY FROM BIDDERS</b>   | <b>HOSPITAL REPLY</b>   |
| <b>1</b>         | Since the tenderers have to provide the operational sterilization services in addition to the supply of the equipments to the hospital as part of the tender document, the supply of RADIO OPAQUE dressing gauze and dressing pad/ rolls for the dressing of the patients is essential requirement. These items are to be issued from the CSSD after sterilization . PLEASE CLARIFY, WILL THE HOSPITAL PROVIDE ALL THE DRESSING MATERIAL FOR STERLISATION AND SUPPLY TO THE HOSPITAL AREAS AS IT IS HIGH CONSUMPTION& COST INTENSIVE ITEM | <b>ALL THE DRESSING MATERIAL for the dressing of the patients FOR STERLISATION AND SUPPLY TO THE HOSPITAL IS IN THE SCOPE OF HOSPITAL.</b>                      |
| <b>2</b>         | Secondly , there is mention of Oxygen ,Nitrous oxide and Vacuum in the compressed air details given on page 57 of the tender document where as these services are not at all needed in CSSD. Kindly clarify if this is still to be taken into consideration   | <b>The hospital will provide compressed air of 4 BAR / 7 BAR at one point in the CSSD area. All other requirements of CSSD shall be in the scope of bidder.</b> |
| <b>3</b>         | Thirdly, this kind of wall panelling (Ref.page 56in tender document) is not normally done for CSSD. This however shall increase the costing and hence do you still feel it mandatory. Please clarify  | <b>Tender conditions prevails.</b>  |