

Date : 29.11.2007

Corrigendum-II

IFB No. HSCC/PUR/GT-Mumbai/2007 dt. 05.11.2007.

Please refer to Corrigendum-I dt. 28.11.2007 displayed on HSCC website. Please note that Corrigendum-I stands cancelled w.r.t. tender specifications and it is informed to all concerned to follow tender specifications as per original bid document.

Corrigendum-II supersedes Corrigendum-I as follows :

Tender Specifications

Name of Equipment	Original Tender Specifications	Amended Tender Specifications
MRI	<u>S. No. 3.2.1 (Gradient System)</u> Actively shielded Gradient System with strength of at least 40 mT/m or more with the slew rate of 150 mT/m/sec or more. The slew rate of 150 at 40 mT should be available in each axis independent for overall better duty cycle performance of the gradient.	No amendment. The technical specifications given in original tender technical bid should be observed.
CT Scan	<u>S. No. 3.7.2 (Data Acquisition System)</u> Minimum 32 rows.	No amendment. The technical specifications given in original tender technical bid should be observed.

As informed by Gokuldas Tejpal Hospital, Mumbai, the technical specifications in original bid serve the purpose of state of art of MRI & CT Scan Machines for G.T. Hospital, Mumbai.

Bid Document Fee :

Clause ITB 9 [B] should be read as follows :

Bid Document Fee is Rs. 2500/- for each Schedule. Bid Document Fee paid is non-refundable and the Bid Documents are non-transferable.

All other terms & conditions, specifications, date of submission remains unaltered.

Executive Director (PG-1 & Proc)

