

Date : 28.11.2007

Corrigendum-I

IFB No. HSCC/PUR/GT-Mumbai/2007 dt. 05.11.2007.

Please note the following amendments in the tender specifications for CT Scan & MRI Systems for Gokuldas Tejpal Hospital, Mumbai.

Tender Specifications

Name of Equipment	Original Tender Specifications	Amended Tender Specifications
MRI	<u>S. No. 3.2.1 (Gradient System)</u> Actively shielded Gradient System with strength of at least 40 mT/m or more with the slew rate of 150 mT/m/sec or more. The slew rate of 150 at 40 mT should be available in each axis independent for overall better duty cycle performance of the gradient.	Maximum gradient strength with maximum slew rate in all three axis should be provided. Two separate head coil & neuro-vascular coil are needed.
CT Scan	<u>S. No. 2.1 (Operational Requirements)</u> Scanner for high resolution all body scanning including cardiac, vascular applications. The system should be capable of acquiring 64 slices per rotation.	64 rows of detectors should be provided.

Bid Document Fee :

Clause ITB 9 [B] should be read as follows :

Bid Document Fee is Rs. 2500/- for each Schedule. Bid Document Fee paid is non-refundable and the Bid Documents are non-transferable.

All other terms & conditions, specifications, date of submission remains unaltered.

Executive Director (PG-1 & Proc)