

Amendment-I**Date :10.06.2025****Subject: Procurement, Supply, Installation & Training of ALS & BLS Ambulance for Kabul, Afghanistan**

Tender/ NIT No.HSCC/PUR/MEA-AFGHANISTAN/AMBULANCE/2025/112 Dated 30.05.2025

This amendment shall form integral part of tender document and should be submitted duly signed & stamped along with the tender/bid.

Reply to Pre Bid Queries raised by bidders through Email during pre -bid meeting held on 04.06.2025 at HSCC, Head Office, Noida

Sr. No.	Clause No.	Bidders Queries	Reply/ Amended as
1	Medical Equipment-Item No.9 Cardiac Monitor	As there is always space crunch in Ambulance Item No. 2 (Defibrillator cum Multi Para Monitor Patient Monitor) has all functions of Cardiac Monitor & this will unnecessarily increase the cost of ALS Ambulance You are requested to delete Cardiac Monitor.	No Change, tender terms & conditions remain unchanged
2	Medical Equipment-Item No.12 ECG Machine	This part is already covered in Defibrillator cum Multi Para Monitor Patient Monitor, why to block space in ambulance. This item may kindly be deleted.	No Change, tender terms & conditions remain unchanged
3	Medical Equipment-Item No.13 Pulse Oximeter	Pulse Oximeter is also the part of Defibrillator cum Monitor. Item No. 2 (Defibrillator cum Multi Para Monitor Patient Monitor) has all functions of Pulse Oximeter & this will unnecessarily increase the cost of ALS Ambulance You are requested to delete the same.	No Change, tender terms & conditions remain unchanged
4	Medical Equipment-Item No.15 AED	AED is also the part of Defibrillator cum Monitor Item No. 2 (Defibrillator cum Multi Para Monitor Patient Monitor) has all functions of AED. This item may kindly be deleted.	No Change, tender terms & conditions remain unchanged
5	Ambulance Origin and Importation	a. Can the ALS and BLS ambulances be directly imported from a foreign country (e.g., UAE, Turkey, China) and supplied to Kabul, or is procurement from Indian manufacturers/OEMs mandatory? b. If direct foreign procurement is permitted, are there any customs/import requirements or exemptions applicable under the agreement with MEA for delivery in Kabul?	No Change, tender terms & conditions remain unchanged
6	Currency of Quotation	a. The tender mandates quoting in Indian Rupees. If the ambulances are sourced from outside India, can the cost component (e.g., ex-works + international freight + customs + inland transit) be reimbursed or supported in a foreign currency payment to the OEM? b. In case of imports, what proof of procurement and value in INR is acceptable (e.g., Proforma Invoice, AD Bank Certificate)?	No Change, tender terms & conditions remain unchanged
7	Make in India Policy and Restrictions	a. As per DPIIT's "Make in India" policy reference in the tender, are non-Indian manufactured ambulances disqualified if they are fully compliant with the technical specifications and delivered DDP Kabul? b. Is preference given to Indian OEMs over foreign OEMs, even if both are technically responsive and commercially competitive?	No Change, tender terms & conditions remain unchanged
8	Pre-Dispatch Inspection	a. Will HSCC/MEA team travel for inspection in foreign countries if the vehicle is manufactured outside India? b. Can internationally recognized third-party inspection (e.g., SGS, TUV, Bureau Veritas) be accepted for such foreign-origin vehicles?	No Change, tender terms & conditions remain unchanged
9	Warranty and After-Sales Support in Kabul	a. For foreign-manufactured ambulances, is it acceptable to maintain an authorized service partner or workshop in Kabul for warranty compliance? b. Is an Indian after-sales network mandatory even for direct-to-Afghanistan delivery?	No Change, tender terms & conditions remain unchanged

Sr. No.	Clause No.	Bidders Queries	Reply/ Amended as
10	2.11 Defibrillator cum Multipara Monitor patient Monitor: Should have fast charging time, charging 200J in 5 seconds or less on both mains and battery.	Should have fast charging time, charging 200J in 6 seconds or less on both mains and battery.	No Change, tender terms & conditions remain unchanged
11	2.13 Defibrillator cum Multipara Monitor patient Monitor: Should have advanced wireless CPR audio visual feedback system which can be used in synchronization with defibrillator provides real time feedback from	Should have advanced Wired/wireless CPR audio visual feedback system which can be used in synchronization with defibrillator provides real time feedback from each compression, guides compression for depth, rate and release. It can be used as training assist device	No Change, tender terms & conditions remain unchanged
12	Section – IX, Qualification Criteria g.Manufacturer Authorization to be submitted for vehicle (OEM) and for the major Equipment's (OEM) at the Time of Bid Submission, as per the below list. i.Defibrillator cum Multi Para monitor Patient Monitor ii.Syringe Pump iii.Transport Ventilator iv.Suction Pump (Electronic) v.Trolley Stretcher- with back tilt facility and collapsible wheels for uploading into the trolley vi.Stretcher Scoop vii.Spinal Board viii.Cardiac Monitor However Bidder has to quote all equipment's with Make/Model, Compliance sheet, Catalogues and Certification for all other equipment's and MAF Form at the time of supply.		Section – IX, Qualification Criteria g.Manufacturer Authorization to be submitted for vehicle (OEM) and for the major Equipment's (OEM) at the Time of Bid Submission, as per the below list. i.Defibrillator cum Multi Para monitor Patient Monitor ii.Syringe Pump iii.Transport Ventilator iv.Suction Pump (Electronic) v.Trolley Stretcher- with back tilt facility and collapsible wheels for uploading into the trolley vi.Stretcher Scoop vii.Spinal Board viii.Cardiac Monitor However Bidder has to quote all equipment's with Make/Model, Compliance sheet, Catalogues and Certification for all other equipment's and for remaining items Manufacture Authorization Form to be submitted at the time of supply.

All other terms & conditions remain unchanged.

Prospective bidders are advised to regularly visit through HSCC e-tender portal <https://hsccltd.co.in>, HSCC website <http://www.hsccltd.co.in> & CPP Portal <https://eprocure.gov.in/epublish/app> as corrigendum/amendments etc., if any, will be notified on this portal only and not be published anywhere else.

(- sd -)
General Manager
HSCC (India) Ltd.